

BuffAR pH 7.0 Reference Solution  
StandARd®



Material No.: 0031-04  
Batch No.: 0000084256  
Manufactured Date: 2014/07/03  
Expiration Date: 2016/01/01

## Certificate of Analysis

Test	Specification	Result
pH at 25°C	6.96 – 7.04	6.97

For Laboratory, Research or Manufacturing Use  
Traceable to NIST

Country of Origin: US  
Packaging Site: Paris Mfg Ctr & DC



Phillipsburg, NJ 9001:2008, 14001:2004  
Paris, KY 9001:2008  
Mexico City, Mexico 9001:2008  
Deventer, The Netherlands 9001:2008, 14001:2004, 13485:2003  
Gliwice, Poland 9001:2008, 17025:2005  
Selangor, Malaysia 9001:2008  
Dehradun, India, 9001:2008, 14001:2004, 13485:2003  
Mumbai, India, 9001:2008, 17025:2005  
Panoli, India 9001:2008

Richard M Siberski  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600  
Avantor™ Performance Materials Inc.

3477 Corporate Parkway, Suite #200, Center Valley, PA 18034, U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610