

Material No.: 0062-20  
Batch No.: 0000227248  
Manufactured Date: 2019/04/02  
Expiration Date: 2022/04/01  
Revision No: 1

## Certificate of Analysis

Test	Specification	Result
Particles (3-30µm)	<= 2	1
pH at 25°C	7.35 - 7.45	7.39
Conductivity, mS*	16.75 - 17.75	17.19
Osmolality, mOsm/kg*	337 - 345	338
Sodium (Na), mmol/L	180 - 188	180
Potassium (K), mmol/L	5.50 - 6.00	5.62
Absorption, maxima (265-275nm)	0.180 - 0.210	0.185
Fluorescence Spectrum	Passes Test	PT
Microbiological Limit Test (cfu/100mL)*	<= 5	< 1
Becton-Dickinson C/A Approval	Passes Test	PT

For Laboratory, Research or Manufacturing Use

\*These stability properties will remain within specification through the shelf life of the product.

Storage Conditions: DO NOT FREEZE, storage temperature, 2-30°C

Country of Origin: US

Packaging Site: Paris Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality