

Material No.: 0062-20  
Batch No.: 0000279227  
Manufactured Date: 2021/03/17  
Expiration Date: 2024/03/16  
Revision No: 1

## Certificate of Analysis

| Test                                    | Specification | Result |
|---|---------------|--------|
| Particles (3-30µm)                      | <= 2          | 1      |
| pH at 25°C                              | 7.34 - 7.50   | 7.41   |
| Conductivity, mS*                       | 16.75 - 17.75 | 17.25  |
| Osmolality, mOsm/kg*                    | 335 - 351     | 346    |
| Sodium (Na), mmol/L                     | 180 - 188     | 180    |
| Potassium (K), mmol/L                   | 5.50 - 6.15   | 5.74   |
| Absorption, maxima (265-275nm)          | 0.180 - 0.210 | 0.185  |
| Fluorescence Spectrum                   | Passes Test   | PT     |
| Microbiological Limit Test (cfu/100mL)* | <= 5          | < 1    |
| Becton-Dickinson C/A Approval           | Passes Test   | PT     |

For Laboratory, Research or Manufacturing Use

\*These stability properties will remain within specification through the shelf life of the product.

Storage Conditions: DO NOT FREEZE, storage temperature, 2-30°C

Country of Origin: US

Packaging Site: Paris Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality