

Material No.: 0062-20
Batch No.: 0000279928
Manufactured Date: 2021/03/26
Expiration Date: 2024/03/25
Revision No: 1

Certificate of Analysis

Test	Specification	Result
Particles (3-30µm)	<= 2	2
pH at 25°C	7.34 - 7.50	7.40
Conductivity, mS*	16.75 - 17.75	16.96
Osmolality, mOsm/kg*	335 - 351	344
Sodium (Na), mmol/L	180 - 188	183
Potassium (K), mmol/L	5.50 - 6.15	5.77
Absorption, maxima (265-275nm)	0.180 - 0.210	0.182
Fluorescence Spectrum	Passes Test	PT
Microbiological Limit Test (cfu/100mL)*	<= 5	1
Becton-Dickinson C/A Approval	Passes Test	PT

For Laboratory, Research or Manufacturing Use

*These stability properties will remain within specification through the shelf life of the product.

Storage Conditions: DO NOT FREEZE, storage temperature, 2-30°C

Country of Origin: US

Packaging Site: Paris Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality